

Buffer Solution pH6.86

[해당없음] (자료없음) FW:자료없음

Product code. 1026

Buffer Grade

TESTS

UNIT

SPECIFICATION

Appearance
pH (25°C)

-
-

Clear
6.84~6.88

Test Method : JIS K 8001

Exp. Date : 3 years after Mfg. Date

DUKSAN PURE CHEMICALS CO. LTD

Ansan | 53, Sinwon-ro 133beon-gil, Danwon-gu, Ansan-si, Gyeonggi-do, Korea
JinCheon | 13-37, Sansusandan 1-ro, Deoksan-eup, Jincheon-gun,
Chungcheongbuk-do, Korea
TEL : +82-31-495-6886 FAX : +82-31-495-5335
World Wide Web : WWW.DUKSAN.KR



Quality Assurance Director