REAGENTS
DUKSAN

Certificate of Analysis

Water for gradient

[7732-18-5] (H2O) FW:18

Product code. 119 **HPLC Grade**

TESTS	UNIT	SPECIFICATION
UV Absorbance 200 nm	-	Max. 0.01
UV Absorbance 250-400 nm	-	Max. 0.005
Gradient Elution test	-	-
- Largest eluted peak(254nm)	mAU	Max. 1
- Largest eluted peak(210nm)	mAU	Max. 5
Fluroescence Trace Impurities	-	-
- at 450nm Emission	ppb	Max. 0.1
- at Emission Maximum for impurities	ppb	Max. 0.2
Residue after evaporation	ppm	Max. 2
Filtered by 0.2um filter	-	-

Test Method : ACS 11th Edition

Exp. Date : 2 years after Mfg. Date

DUKSAN PURE CHEMICALS CO. LTD

Ansan | 53, Sinwon-ro 133beon-gil, Danwon-gu, Ansan-si, Gyeonggi-do, Korea
JinCheon | 13-37, Sansusandan 1-ro, Deoksan-eup, Jincheon-gun, Chungcheongbuk-do, Korea
TEL : +82-31-495-6886 FAX : +82-31-495-5335
World Wide Web : WWW.DUKSAN.KR

Quality Assurance Director