REAGENTS
DUKSAN

Certificate of Analysis

Nitric Acid 69%

[7697-37-2] (HNO3) FW:63.01

Product code. 4775

GR/AR/ACS Grade

TESTS	UNIT	SPECIFICATION
Appearance	-	Colorless liquid,reddish-yellow color with light
Identification	-	Generate yellowish-red color vapour with copper strip
Assay	%	69.0~70.0
Ignition residue (as Sulfate)	ppm	Max. 3
Chloride (Cl)	ppm	Max. 0.3
Sulfate (SO4)	ppm	Max. 0.5
Copper (Cu)	ppm	Max. 0.1
Lead (Pb)	ppm	Max. 0.1
Arsenic (As)	ppm	Max. 0.01
Iron (Fe)	ppm	Max. 0.2

Test Method : JIS/AR/ACS

Exp. Date : 5 years after Mfg. Date

DUKSAN PURE CHEMICALS CO. LTD

Ansan | 53, Sinwon-ro 133beon-gil, Danwon-gu, Ansan-si, Gyeonggi-do, Korea
JinCheon | 13-37, Sansusandan 1-ro, Deoksan-eup, Jincheon-gun, Chungcheongbuk-do, Korea
TEL : +82-31-495-6886 FAX : +82-31-495-5335
World Wide Web : WWW.DUKSAN.KR

Quality Assurance Director